

Interlock scheme at USA15 (¼ partition)

Each group of 4 fingers is served by an independent current loop. All loops are coupled to the DSS alarm through the Interlock Board.

The HV crate has no internal Isource => in series with AUX board 4.

